

Piraeus, 30 March 2022

Piraeus Port Authority S.A.

Procurement Department

10, Akti Miaouli,

GR 185 38,

Piraeus, Greece.

Tel: +30 210 4550186

E-mail: procurement@olp.gr

SUBJECT: Clarifications to the interested parties, regarding the tender for the procurement of measuring system for measuring particulate matter 2.5 concentration analyzer pm2.5.

Interested parties are kindly requested to refer to all clarifications published provided by PPA, in relation to the subject tender.

Question 1:

“ We are interested to participate in this tender but need some technical clarification on the data transmission, ie if it will be accomplished via the station's data logger or independently, and over which data highway (cabled Ethernet, 3G etc.). Please advise us of a knowledgeable colleague that could advise us on that.

Also would it be possible to visit your AQM station within this week”.

Answer 1

Επί της ερώτησης που έχει τεθεί σημειώνεται ότι στις τεχνικές προδιαγραφές αναφέρονται τα ακόλουθα:

- Θύρες: Να διαθέτει ethernet, RS 232 και αναλογική έξοδο για σύνδεση με σύστημα data logging. Να υπάρχει ενσωματωμένη μνήμη αποθήκευσης των μετρήσεων και των πιθανών προβλημάτων του αναλυτή (διαγνωστικά) διάρκειας ενός μήνα τουλάχιστον.
- Να έχει τη δυνατότητα προγραμματισμού του ρυθμού μέτρησης και αποστολής των μετρήσεων.
- Εγκατάσταση του λογισμικού τελικού χρήστη (user interface) σε τέσσερεις (4) χρήστες
- Ως προς την επίσκεψη στον σταθμό θα μπορούσε να υπάρξει απευθείας επικοινωνία με τον ενδιαφερόμενο για να κανονισθεί.